


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-2-PS-T-N*

S.N.: *1866*

Date: *MAY 28 84*



- a) Output signal amplitude:  
*0 TO -350V.*
- b) Pulse width:  
*0 TO 100 NSEC.*
- c) Rise time:  
*≤ 1.5 NSEC*
- d) Fall time:  
*TO 7, 10 NSEC*  
*≤ 3 NSEC.*
- e) PRF:  
*0 TO 5 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*320 VOLTS/DIV.*  
*10 NSEC/DIV.*  
*4 kHz.*

*110/220V*  
*50-60 Hz.*