


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-AV-1-C-CA-EN-PN-M-05*

S.N.: *1872*

Date: *JUNE 26 84*



~~100~~  
*100 VOLTS/DIV*  
*1.0 nSEC/DIV*  
*.500 KHz*

- a) Output signal amplitude:  
*0 TO ±10 VOLTS*
- b) Pulse width:  
*0 TO 4 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 135 pSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220 V,*  
*50-60 Hz.*