


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PRR-1-PW-C-6A 6W-PN-05*


S.N.: *1874*

Date: *JUNE 25 84*



*P OUT*  
*50 VOLTS/DIV.*  
*1.0 USEC/DIV*

- a) Output signal amplitude:  
*0 TO ±200V*
- b) Pulse width:  
*0.1 to 5.0 usec*
- c) Rise time:  
*≤ 10 nsec*
- d) Fall time:  
*≤ 10 nsec*
- e) PRF:  
*0 TO 1 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:



*N OUT.*  
*50 VOLTS/DIV.*  
*1.0 USEC/DIV*

*110/220V,*  
*50-60 Hz.*