


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-A-C-05-P*

S.N.: *1887*

Date: *JULY 10 84*



*10 VOLTS/DIV.
1.0 NSEC/DIV.
100 KHz*

- a) Output signal amplitude:
0 TO +15V.
- b) Pulse width:
4 TO 100 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC.
- e) PRF: *SEE FIGS 1, 2.*
- f) Jitter, stability:
OK
- g) Prime power:

*110/220V
50-60 Hz*