


PULSE GENERATOR  
PERFORMANCE CHECK

Model: M6-2-P

S.N.: 1838

Date: APRIL 4 84



- a) Output signal amplitude:  
0 TO 4120 VOLTS
- b) Pulse width:  
≤ 2 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 20 KHZ
- f) Jitter, stability:  
O.K.
- g) Prime power:

100 VOLTS/DIV

2.0 NSEC/DIV

10 KHZ

+15 ± 1.0 VOLTS  
150 mA (MAX)