


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-AV-N*
S.N.: *1890*
Date: *AUG 27 84*



- a) Output signal amplitude:
0 TO -15 VOLTS
- b) Pulse width:
2 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

5 NSEC/DIV.
10 VOLTS/DIV.
50 KHz.

+15 ± 1.0 VOLT
300 mA (MAX)