

PULSE GENERATOR

PERFORMANCE CHECK

Model:


A20-5.13-PS-SL-1

S.N.:

1891

Date:

APR 6 84



- a) Output signal amplitude:
0 TO +30 AMPS TO
3 017MS
- b) Pulse width:
0.2 TO 4.6 USEC
- c) Rise time:
 ≤ 20 NSEC
- d) Fall time:
 ≤ 20 NSEC
- e) PRF:
0 TO 300 HZ.
- f) Jitter, stability:
OK
- g) Prime power:
110/220V
50-60 HZ

6.5 AMPS / DIU

200 NSEC / DIU

300 HZ.

RL = 3 Ω