


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MVL-2-C-DNT-M-OS*

S.N.: *1844*

Date: *JUNE 21 1984*



*≈ 320 VOLTS/DIV  
10 NSEC/DIV  
1.0 KHz*

- a) Output signal amplitude:  
*0 TO ± 350 VOLTS*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 1.5 NSEC*
- d) Fall time:  
*≤ 3 NSEC*
- e) PRF:  
*0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220 V  
50-60 Hz*