


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-HV3-C-PN

S.N.: 1898

Date: JULY 27, 1984



- d) Output signal amplitude:
0 TO $\pm 40V$...
- e) Pulse width:
0 TO 2 NSEC
- f) Rise time:
 ≤ 150 PSEC
- g) Fall time:
 ≤ 250 PSEC
- h) PRF:
0 TO 100 KHZ
- i) Jitter, stability:
OK
- j) Prime power:

32 VOLTS/DIV
1.0 NSEC/DIV
100 KHZ

110/220V
50-60 KHZ