

PULSE GENERATOR

PERFORMANCE CHECK

Model:


PVN-1-C-PN-M

S.N.:

1901

Date:

AUG 20 84



a) Output signal amplitude:

0 TO 5 VOLTS

b) Pulse width:

0.2 TO 1.0 USEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 135 PSEC

e) PRF:

50 TO 250 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V,

50-60 Hz

3.2 VOLTS/DIV
1.0 USEC/DIV
(200 MHz)