


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-C-N*

S.N.: *1902*

Date: *JULY 9 84*



-) Output signal amplitude:
0 TO 50 V
-) Pulse width:
2 TO 100 NSEC
-) Rise time:
≤ 300 PSEC
-) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
110/220 V

32 VOLTS/DIV

5 NSEC/DIV

10 KHZ

50-60 Hz