


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1S-C-PN-OS-M

S.N.: 1912

Date: JULY 27, 1984



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
 ≤ 60 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
500 PSEC/DIV
100 KHz

110/220V,
50-60 Hz