


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-2-W-P-CVA-EW*
S.N.: *1915*
Date: *Aug 24 1984*



- a) Output signal amplitude:
0 TO +10 VOLTS
- b) Pulse width:
20 TO 500 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 200 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
5.0 NSEC/DIV
50 KHz

+ 24 ± 1.0 VOLTS
300 mA (MAX)