


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-HV3-P

S.N.: 1919

Date: JULY 27, 1984.



a) Output signal amplitude:

0 TO +80V

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

≤ 150 PSEC

d) Fall time:

≤ 250 PSEC

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 \pm 1.0 VOLTS

300 mA (MAX)

32 VOLTS/DIV
1.0 NSEC/DIV
400 KHZ