


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-PW-C-PN


S.N.: 1920

Date: SEPT. 21, 1984



Ⓐ 50 VOLTS/DIV.
100 NSEC/DIV
P OUT

- a) Output signal amplitude:
0 TO ± 200 VOLTS
- b) Pulse width:
50 NSEC TO 1.0 USEC
- c) Rise time:
 ≤ 10 NSEC
- d) Fall time:
 ≤ 10 NSEC
- e) PRF:
0 TO 100 KHz
(0.5% MAX. DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
110/220 V,
50-60 Hz



Ⓑ 50 VOLTS/DIV.
100 NSEC/DIV
N OUT.