


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-05-P*
S.N.: *1929*
Date: *AUG 10 84*



- a) Output signal amplitude:
0 TO + 40 VOLTS
- b) Pulse width:
2 TO 100 NSEC.
- c) Rise time:
≤ 500 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
SEE FIGS 1, 2
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
2.0 NSEC/DIV
100 KHz.

+15 ± 1.0 V
200 MA (MAX)