



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVR-A-1-P*
S.N.: *1437*
Date: *Aug 10 1984*



- a) Output signal amplitude:
0 TO +200 V TO 50 Ω
- b) Pulse width:
50 nSEC TO 1 μ SEC
- c) Rise time:
 \leq 10 nSEC
- d) Fall time:
 \leq 10 nSEC
- e) PRF: *0 TO 100 kHz*
0.5% MAX DUTY CYCLE
- f) Jitter, stability:
OK
- g) Prime power:
+24 \pm 1.0 VOUTS
350 mA (MAX)

Ⓐ *50 V/DIV*
2.0 μ SEC/DIV
R_L = 50 Ω
PRF = 400 Hz



Ⓑ *AS Ⓐ BUT*
0.2 μ SEC/DIV

[Signature]