


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-A-PS-P*

S.N.: *1956*

Date: *JUNE 28 84*



- a) Output signal amplitude:  
*0 TO +50 V.*
- b) Pulse width:  
*0 TO 100 NSEC*
- c) Rise time:  
*≤ 300 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*0 TO 20 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*32 VOLTS/DIV.  
10 NSEC/DIV.  
10 KHz.*

*(R<sub>c</sub> = 50 OHMS)*

*110/220 V  
50-60 Hz*