


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-P-M

S.N.: 1964

Date: OCT 9/84



- a) Output signal amplitude:
0 TO + 350V.
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

320 VOLTS/DIV
10 NSEC/DIV
1 KHZ.

110/220V
50-60 Hz