


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-1S-C-P-0S

S.N.: 1972

Date: JULY 27, 1984



a) Output signal amplitude:

0 TO +5V

b) Pulse width:

0 TO 4 μ S

c) Rise time:

\leq 60 PSEZ

d) Fall time:

\leq 130 PSEZ

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz

5 VOLTS/DIV
500 PSEZ/DIV
100 KHz