


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-HV3-C-OS-PN-M

S.N.: 1974

Date: JULY 27, 1984



- a) Output signal amplitude:
0 TO ± 40 VOLTS
- b) Pulse width:
0 TO 2 NSEC
- c) Rise time:
 ≤ 150 PSEC
- d) Fall time:
 ≤ 250 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

110/220 V,
50-60 Hz

32 VOLTS/DIV
1.0 NSEC/DIV
400 KHz.