


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-HV3-PS-PN

S.N.: 1975

Date: JULY 27, 1984



a) Output signal amplitude:

0 TO  $\pm 40$  VOLTS

b) Pulse width:

0 TO 2  $\mu$ SEC

c) Rise time:

$\leq 150$  PSEC

d) Fall time:

$\leq 250$  NSEC

e) PRF:

0 TO 1 IMP/SEC

f) Jitter, stability:

OK

g) Prime power:

32 VOLTS/DIV  
1.0  $\mu$ SEC/DIV  
400 KHZ

110/220 V,  
50-60 HZ.