


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVT-V-1403-C-PN-05-M*

S.N.: *1999*

Date: *SEPT 10 1984*



- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width:
3 TO 100 NSEC
- c) Rise time:
 ≤ 200 PSEC
- d) Fall time:
 ≤ 2 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

110/220 V

50-60 HZ

100 VOLTS/DIV

5 NSEC/DIV

[Signature]