


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVL-AV-1-C-PN* *-05-MA-W-T*

S.N.: *2000*

Date: *SEPT 10 1984*



- a) Output signal amplitude:
0 TO ±100 VOLTS
- b) Pulse width:
5 TO 400 NSEC
- c) Rise time:
≤ 1 NSEC
- d) Fall time:
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

100 VOLTS/DIV
5 NSEC/DIV

110/220V
50-60 Hz