

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-PS-PN-M

S.N.: 2005

Date: Nov. 1, 1984.

a) Output signal amplitude:

0 TO  $\pm 350$  VOLTS

b) Pulse width:

0 TO 100 NSEC.

c) Rise time:

$\approx 1.5$  NSEC

d) Fall time

$\approx 3$  NSEC

e) PRF:

0 TO 5 KHZ


f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz



$\approx 320$  VOLTS/DIV

10 NSEC/DIV

1.0 KHZ