


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-3-P-1D*
S.N.: *2007*
Date: *Aug 15 84*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
*2 TO 15 nSEC**
- c) Rise time:
≤ 150 pSEC
- d) Fall time:
≤ 600 pSEC
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+28V, 350 mA max

10 VOLTS/DIV
5 nSEC/DIV
1.0 MHz

** WILL PROVIDE
20 nSEC AT
1.0 MHz.*