


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-M2-N

S.N.: 2008

Date: SEPT. 19, 1984



a) Output signal amplitude:

≥ -60 VOLTS

b) Pulse width:

≤ 2 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

$+135 \pm 5$ V

5 mA (MAX.)

32 VOLTS/DIV

2 NSEC/DIV

500 KHZ