


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-HV3-C-P

S.N.: 2012

Date: OCT 12, 1984



- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width:
0.4 TO 2.5 NSEC
- c) Rise time:
≤ 150 PSEC
- d) Fall time:
≤ 250 PSEC
- e) PRF:
0 TO 1 MHZ
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
1.0 NSEC/DIV
100 KHZ

110/220 V
50-60 HZ