


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-AV-HV3-N

S.N.: 2013

Date: Nov. 5, 1984



- d) Output signal amplitude:
0 to 40 V
- e) Pulse width:
0 to 2 nsec
- f) Rise time:
 ≤ 150 psec
- g) Fall time:
 ≤ 250 psec
- h) PRF:
0 to 1 MHz
- i) Jitter, stability:
OK
- j) Prime power:

≈ 32 VOLT/Div
1.0 nsec/Div
500 kHz

+15 \pm 1.0 V
200 mA (max)
