


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AYP - AV - HV3 - N

S.N.: 2014

Date: NOV. 5, 1984



- a) Output signal amplitude:  
0 TO -40 V
- b) Pulse width:  
0 TO 2 NSEC
- c) Rise time:  
≤ 150 PSEC
- d) Fall time:  
≤ 250 PSEC
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

≈ 32 VOLTS/DIV  
100 NSEC/DIV  
500 KHz

+15 ± 1.0 VOLTS  
200 mA (MAX)