


PULSE GENERATOR

PERFORMANCE CHECK

Model: *ANMN-1-P*

S.N.: *2013*

Date: *AUG 28 1984*



- a) Output signal amplitude:  
*0 TO +5 VOLTS*
- b) Pulse width:  
*0.2 TO 6 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 50 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+24 ± 1.0 VOLT*  
*400 mA (MAX)*

*5 VOLTS/DIV*  
*5 NS/DIV*  
*40 MHz*

*[Signature]*