


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AN-1 -N.*

S.N.: *2020*

Date: *AUG 28 1984*



- a) Output signal amplitude:
0 TO -5V
- b) Pulse width:
0.2 TO 6 NSEC
- c) Rise time:
≤ 100 PSEC.
- d) Fall time:
≤ 200 NSEC.
- e) PRF:
0 TO 50 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
5 NSEC/DIV
40 MHz.

+24 ± 1.0 VOLT
400 mA (MAX)