


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AP-AV-1-N*

S.N.: *2023*

Date: *Jul 15 84*



- a) Output signal amplitude:  
*0 to 10V*
- b) Pulse width:  
*0.2 to 4.0 nsec*
- c) Rise time:  
*≤ 100 psec*
- d) Fall time:  
*≤ 135 psec*
- e) PRF:  
*0 to 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*10 VOLTS/DIV  
1.0 nSEC/DIV  
400 kHz*

*+15 ± 1.0 VOLT  
300 mA (MAX)  
*[Signature]**