

PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-W-C-S1

S.N.: 2024

Date: OCT. 25/84

a) Output signal amplitude:

P: 0 TO +10V (10V)
N: 0 TO -20V (10V)

b) Pulse width:

P: 25 TO 100 μ SEC

N: 25 TO 100 μ SEC

c) Rise time:

\leq 0.1 μ SEC

d) Fall time:

\leq 0.1 μ SEC

e) PRF:

0 TO 1 KHz


f) Jitter, stability:

OK

g) Prime power:

110/220V

50-60 Hz



5 VOLTS/DIV

10 μ SEC/DIV

1 KHz