


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-W-C-S2

S.N.: 2025

Date: OCT 25/84



- a) Output signal amplitude:
P: 0 TO +20 V TO 10Ω
N: 0 TO -20 V TO 10Ω
- b) Pulse width:
P: 1 TO 10 μSEC
N: 1 TO 10 μSEC
- c) Rise time:
≤ 5 nSEC
- d) Fall time:
≤ 5 nSEC
- e) PRF:
0 TO 1.0 KHz
- f) Jitter, stability:
OK.
- g) Prime power:
110/220 V
50-60 Hz.

10 V/DIV
0.5 μSEC/DIV
1 KHz