


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HVI-P

S.N.: 2031

Date: SEPT 20/84



- a) Output signal amplitude:  
0 TO + 10.0 VOLTS
- b) Pulse width:  
≤ 1 NSEC
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 100 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

100 VOLTS/DIV

1.0 NSEC/DIV

50 KHz

+15 ± 1.0 V

200 mA (MAX)

*[Handwritten Signature]*