


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVM-2-P

S.N.: 2036

Date: OCT 12, 1984



a) Output signal amplitude:

0 TO +15V

b) Pulse width:

0.2 TO 2.0 NS

c) Rise time:

≤ 100 PS

d) Fall time:

≤ 135 PS

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 ± 1.0 VDC
558 mA (MAX)

10 VOLTS/DIV

1.0 NS/DIV

5 MHz