


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVL-2-C-T-P-OS-M

S.N.: 2037

Date: OCT. 30/84



- a) Output signal amplitude:
0 TO +350V
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 1.5 NSEC
- d) Fall time: (TO 10 NSEC)
≤ 3 NSEC
- e) PRF:
0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

≈ 320 V/DIV
10 NSEC/DIV

110/220V
50-60 Hz