


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-PS-T-N

S.N.: 2040

Date: OCT. 30/84



- a) Output signal amplitude:
0 TO -350 VOLTS
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
1.5 TO 10 NSEC
- d) Fall time:
3 TO 10 NSEC
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
110/220 V

≈ 320 VOLTS / DIV
10 NSEC / DIV
1 KHZ

50 - 80 HZ

[Signature]