


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1
S.N.: 2043
Date: Nov 1, 1984



- a) Output signal amplitude: +25 TO +100 AMPS
- b) Pulse width: 25 NSEC
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1 KHz
- f) Jitter, stability: OK
- g) Prime power:

≈ 65 AMPS/DIV
10 NSEC/DIV
1.0 KHz

+15 ± 1.0 VOLTS
200 mA (MAX)

AA