


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVMN-1-PS-PN-M

S.N.: 2044

Date: DEC 5, 1984



- ) Output signal amplitude:  
0 TO  $\pm 5V$
- ) Pulse width:  
0 TO 6  $\mu SEC$
- ) Rise time:  
 $\leq 100$   $\mu SEC$
- ) Fall time:  
 $\leq 135$   $\mu SEC$
- ) PRF:  
0 TO 50  $\mu mHz$
- ) Jitter, stability:  
OK
- ) Prime power:

5 VOLT/DIV

1.0  $\mu SEC/DIV$

25  $\mu mHz$

110/220V

50-60 Hz