


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-P

S.N.: 2057

Date: SEPT 19, 1984



a) Output signal amplitude:

0 TO +30 V

b) Pulse width:

5 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

150 mA (MAX)

3.2 VOLTS/DIV

1.0 NSEC/DIV

500 KHz