


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-1-PW-C-P

S.N.: 2072

Date: OCT 26/84



- a) Output signal amplitude:
0 to 7200 V.
- b) Pulse width:
75 nsec to 5 μ sec.
- c) Rise time:
 ≤ 10 nsec
- d) Fall time:
 ≤ 10 nsec
- e) PRF:
0 to 1 kHz.
- f) Jitter, stability:
OK
- g) Prime power:

50 V/DIV
1.0 μ SEC/DIV
1 kHz.

110/220 V
50-60 Hz.