


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-3-PW-PS-N

S.N.: 2073

Date: OCT 26/84



a) Output signal amplitude:

0 TO 200V

b) Pulse width:

0.1 TO 100 USEC.

c) Rise time:

≤ 10 NSEC

d) Fall time:

≤ 10 NSEC

e) PRF:

0 TO 5KHz (MIN PW)

0 TO 50 Hz (MAX PW)

Jitter, stability:

OK

f) Prime power:

110/220V,

50-60 Hz.

50 VOLTS/DIV

70 USEC/DIV

50 Hz.

[Signature]