


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVL-MA-1-P*

S.N.: *2080*

Date: *OCT 19 84*



*100 VOLTS/DIV
2 NSER/DIV*

a) Output signal amplitude:

> +100 V

b) Pulse width:

≈ 6 NSER

c) Rise time:

≤ 700 PSEZ

d) Fall time:

≤ 700 PSEZ

e) PRF:

0 TO 10 KHz

f) Jitter, stability:

OK

g) Prime power:

+ 10.8 TO + 13.2 VOLT?

I_s ≤ 75 mA TYP

PRF = 250 Hz.