


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVL-MA-1-P*

S.N.: *2021*

Date: *OCT 1984*



- a) Output signal amplitude:
> +100 V
- b) Pulse width:
≈ 6 NSEC
- c) Rise time:
≈ 700 PSEC
- d) Fall time:
≈ 700 PSEC
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*100 VOLTS/DIV
2 NSEC/DIV*

*+10.8 TO +13.2 VOLTS
I_s ≈ 75 mA P_{con}
PRF = 256 Hz*