


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-NA-2-T

S.N.: 2086

Date: DEC 20 84



- a) Output signal amplitude:  
     $\geq +80V$
- b) Pulse width:  
    6 nSEC
- c) Rise time:  
     $\leq 1$  nSEC
- d) Fall time:  
     $\leq 1$  nSEC
- e) PRF:  
    0 TO 12 KHz
- f) Jitter, stability:  
    OK
- g) Prime power:

+10.8 TO +13.2 V.  
75 mA (MAX)

100 VOLTS/DIV  
5 nSEC/DIV  
5 KHz