


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVK-A-C-P

S.N.: 2092

Date: SEPT 20, 1984



10 VOLTS/DIV  
1.0 NSEC/DIV  
500 KHz

- a) Output signal amplitude:  
0 TO +15 VOLTS
- b) Pulse width:  
4 TO 100 NSEC
- c) Rise time:  
 $\leq 100$  PSEC
- d) Fall time:  
 $\leq 2$  NSEC
- e) PRF:  
SEE FIGS 1, 2.
- f) Jitter, stability:  
OK
- g) Prime power:  
110/220 V  
50-60 Hz