


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-S-C-B1

S.N.: 2093

Date: DEC 13, 1984



10 VERT/DIV  
1.0 NSEC/DIV  
1.00 KHz

- a) Output signal amplitude:  
0 TO +15V
- b) Pulse width:  
≤ 500 PSEC.
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
5 Hz TO 1.0 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

110/220V

50-60 Hz