


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-13-2-PS-P*

S.N.: *2098*

Date: *DEC 10 84*



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
20 TO 200 NS
- c) Rise time:
≤ 3 NS
- d) Fall time:
≤ 3 NS
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

20 VOLTS/DIV
100 NS/DIV
50 KHZ
R_L = 50 Ω

120/240 V
50-60 HZ

CANTON UNIT
SHIPPED WITH CANTON
IN 120 V POSITION